

Anthony DeGraffenreid, Chemistry

University: Truman State University

Year in School: Junior

Hometown: Eldon, MO

Faculty Mentor: Dr. David Robertson, MU Research Reactor


Funding Source: NSF-REU/NIH Program in Radiochemistry

Analysis of fluorine via $^{19}\text{F}(n, \gamma)^{20}\text{F}$ decay in the presence of a Na interferant using NAA

A. DeGraffenreid, V. Spate, S. Crane, and D. Robertson

Society's concern with the development of innovative diagnostic tools in medicine has prompted a great deal of interest in research. An important area of research includes the development of molecular imaging agents. These agents hold promise for noninvasive in vivo imaging, quantification, and monitoring of important biomarkers for several diseases including atherosclerosis and cancer. This area of research has benefited greatly from instrumental neutron activation analysis. INAA is an analytical technique that is useful for performing quantitative multi-elemental analysis. NAA is superior to other forms of analysis of its ability to simultaneously analyze several elements in a sample and it offers high sensitivity; part per million to low part per billion level. The imaging agent of interest in this project is a lipid encapsulated, liquid perfluorocarbon nanoparticle directly coupled to a selective $\alpha_v\beta_3$ -integrin ligand. The nanoparticle also contains the paramagnetic contrast agent gadolinium linked to the nanoparticle as Gd-DTPA-bis-oleate.¹ Utilizing INAA, analysis of this molecular imaging agent for its concentration in biological tissue, specifically, rabbit aortas via a $^{19}\text{F}(n, \gamma)^{20}\text{F}$ reaction. Measurement of fluorine in tissue is particularly difficult due to fluorine's short half-life ($t_{1/2} = 11\text{s}$) and sodium's interference via a fast neutron reaction. INAA will also be employed to quantitatively determine the concentration of gadolinium in biological tissues. These measurements will allow for the comparison of Gd to F ratio pre-injection to the Gd to F ratio post-injection. Efficiency of Gd's arrival at the target location will be determined by the comparison of these ratios.

Molecular Imaging Agent



¹Ref. see footnote

¹ Characterization and biodistribution of a novel MRI molecular imaging agent by neutron activation analysis

J. D. Robertson,^{1*} S. B. Crane,² S. A. Wickline,³ G. M. Lanza³

¹ University of Missouri Research Reactor and Department of Chemistry, University of Missouri, Columbia, MO 65211, USA

² University of Missouri Research Reactor, University of Missouri, Columbia, MO 65211, USA

³ Department of Cardiology, Washington University Medical School, St. Louis, MO 63110, USA